



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/037,306	12/31/2001	Simon M. Furnish	InFraReDx-12	4514
26161	7590	06/28/2005	EXAMINER	
FISH & RICHARDSON PC 225 FRANKLIN ST BOSTON, MA 02110			JUNG, WILLIAM C	
			ART UNIT	PAPER NUMBER
			3737	

DATE MAILED: 06/28/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

10/037,306

Applicant(s)

FURNISH, SIMON M.

Examiner

William Jung

Art Unit

3737

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 22 June 2005.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-44 and 46-48 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-8, 13-25, 30-37, 39-44 and 46-48 is/are rejected.
- 7) ☒ Claim(s) 9-12, 26-29, and 38 is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. _____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- * See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____.
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____.
- 5) ☐ Notice of Informal Patent Application (PTO-152)
- 6) ☐ Other: _____.

DETAILED ACTION

Response to Arguments

1. Applicant's arguments with respect to claims 1-8, 13-25, 30-37, 39-44, and 46-48 have been considered but are moot in view of the new ground(s) of rejection.

Claim Rejections - 35 USC § 102

2. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

3. Claims 1-3, 20, 24, 25, 33, 39-41, and 44 rejected under 35 U.S.C. 102(e) as being anticipated by *Boppart et al* (US 6,485,413).

Boppart et al anticipate all claimed features in claims 1-3, 20, 24, 25, 33, 39-41, and 44.

Claims 1, 20, 24, 25, 33, 39-41, and 44: Boppart et al disclose a method and apparatus to detect a structure beneath a wall of a body lumen or cavity using a catheter comprising an elongated housing rotatably supported on a catheter sheath with the housing supporting a first and second reflective surfaces with first reflective surface is coupled to delivery fiber and second reflective surface is coupled to collection fiber (col. 2, lines 40-49; col. 2, lines 62-65; col. 4, lines 39-64; col. 7, lines 9-41). Although, the reflective surfaces are not spaced apart, however, they are functionally equivalent to two independent reflective surfaces.

Claims 2 and 3: Boppart et al illustrate in figure 8b where the reflective surface may be a rotating prism that redirects the beam.

Claim Rejections - 35 USC § 103

4. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

5. 4-8, 13-19, 21-23, 30-32, 34-37, 42, 43, and 46-48 are rejected under 35 U.S.C. 103(a) as being unpatentable over *Boppart et al* as applied to claims 1, 2, 20, 33, 39, and 41 above, and further in view of *Winston* (US 5,916,210).

Boppart et al substantially disclose all claimed features in claims 4-8, 13-19, 21-23, 30-32, 34-37, 42, 43, and 46-48. However, Boppart et al do not specifically disclose the following features in which Winston's teaching supplements detailed structure of a similar optical coherent tomography (OCT) system.

Claims 4 and 5: Winston discloses of slot having shoulder or fitting sleeve over the rotating shaft to guide and align the reflective surface and allow the catheter to be manipulated into the sheath (col. 4, lines 23-34).

Claims 6-8, 34-37, 46, and 48: Winston discloses of the catheter described above with the reflector fitting into a pocket like flange where the mirror reflector is fixedly attached to the pocket where the mirror or reflectors redirects the light beams (col. 5, line 65 – col.6, line 11).

Claims 13-15 and 17: Winston's catheter disclosure includes reflective surface being unitary, within the catheter housing described having proximal end and distal ends with energy collecting fiber and slot to align the reflective surface (col. 4, lines 24-57).

Art Unit: 3737

Claim 19: Winston's catheter includes reflective surface enclosed in a catheter tip (col. 5, lines 59-64).

Claims 16, 18, 19, 21-23, and 43: Winston shows in figures 4, 5, and 12a, that the reflective surfaces are in non-parallel position with one being larger than the other and the reflective surface is curvilinear.

Claims 30 and 47: Winston discloses of optical fiber 30 being enclosed within the catheter housing.

Claim 31: Winston discloses that the light delivery fibers are diametrically opposed to the rotation axis to prevent from unbalancing and minimize eccentricity during rotation (col. 9, line 54 – col. 10, line 7).

Claim 32: The reflective surfaces operate independently, therefore the energy delivery is independent.

Therefore, it would have been obvious to one having an ordinary skill in the art at the time the invention was made to apply the teachings of Winston to Boppart et al to improve the OCT imaging of inner walls of lumen.

Allowable Subject Matter

6. Claims 9-12, 26-29, and 38 are objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

Art Unit: 3737

Conclusion


Any inquiry concerning this communication or earlier communications from the examiner should be directed to William Jung, Ph.D. whose telephone number is 571-272-4739. The examiner can normally be reached on Mon-Fri 8:30 AM to 5 PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Brian Casler can be reached on 571-272-4956. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

WJ

June 23, 2005


BRIAN L. CASLER
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 3700